

Mounting the
IS.9150 Cardiod darkfield condenser with
integrated 5 W LED light source and
external 100-240 V mains adapter



1. Unlock hexagon screw and pull out the LED light source




2. Place darkfield condenser



3. Move the condenser up and place the LED light source back into the condenser.



4. Fix LED light source by tightening hexagon screw



5. Device is ready for use.